



UK Declaration of Conformity



Digi International Inc.

9350 Excelsior Blvd., Suite 700, Hopkins, MN 55343 USA

For the following equipment:

Product Details	
Product Name	Capsule SmartLinx Axon 110, 410, 810
Model Number	70001703, 70001704, 70001705
Part Number	50001881-01, 50001869-02, 50001869-03

We declare under our own responsibility that the above product satisfies all the technical regulations applicable to the product within the scope of United Kingdom Regulations:

The above product is in conformity with the following standards or other normative documents:

Regulations	Standards
Electrical Equipment (Safety) Regulations 2016 S.I. 2016 No. 1101	EN 62368-1: 2014/A11:2017 EN 62311: 2008
Electromagnetic Compatibility Regulations 2016 S.I. 2016 No. 1091	EN 301 489-1 V1.9.2 (2011-09) EN 301 489-17 V2.2.1 (2012-09)
Radio Equipment Regulations 2017 S.I. 2017 No. 1206	EN 300 328 V2.1.1 (2016-11) EN 301 893 V2.1.0 (2017-03)
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment (Amendment) Regulations 2021 S.I. 2020 No. 1647	BS EN IEC 63000: 2018
Medical	IEC 60601-1:2005 + Corr. 1:2006 + Corr. 2:2007 + A1:2012 (or IEC 60601-1:2012 reprint) IEC 60601-1-2:2007(ed.3) IEC 60601-1-6:2010(ed.3) + A1:2013 IEC 62366:2007(ed.1) + A1:2014

Minnesota, USA Oct, 4 2021
Digi International Inc.
Erik Reynolds
Director of Engineering
English

UK Representative:
John Pinkney
Digi International Limited
Chancery House
St Nicholas Way
Sutton
SM1 1JB